OQI Report Page 1 of 2

Report for QIS OQI as of 28/06/2022 11:23:13 AM

Report for QIS OQI -

18580 No Title Provided

OQI Details

Status | Closed Approved

Subject When moving the slicprep base and spin basket with collar fron the

centrifuge to the bench of the MPII B platform, the slicprep base dropped from the collar and spin basket on to the MPII B bench and lysis solution splashed onto the bench and possibly into surrounding samples. After the accident, the collar and spin basket was returned to the slicprep base. After consultation with the Analytical PO4, the extraction was continued and all relevant information was entered into

batch and sample audit trails in AUSLAB.

Source of OQI | Clinical Incident
Date I dentified | 10/01/2008

OQI Creator Contact Details

Creator | Breanna GALLAGHER

Organisational Unit/s DNA Analysis

Service/s

Site Location/s | Coopers Plains

Investigator/ Actioner Contact Details

Actioner Allan MCNEVIN

Organisational Unit/ s | Analytical

Service/s

Site Location/s | Coopers Plains

Investigation Details

Investigation Completed Investigation Details Root Cause Type OtherThe above description refers to extraction batch CWIQEXT20071219_03. This resulted from human error.

Preformed By | Quality Information System

Action Details

Action Complete
Title

Action Description each sample and batch audit entries were made to disclose error. Samples were placed onto the appropriate team communications lists in AUSLAB. OQI number entered in UR notes for each sample This OQI will be discussed in the Analytical team meeting 14-1-08. The need to handle the SlicPrep device by the base at all times will be highlighted. Case

OQI Report Page 2 of 2

managing scientists will assess the usability of each sample, ability to resample / resubmite etc.

Task Details

No Tasks found

Follow-up And Approval

Follow-up Status | Accepted

Follow-up Status Comment

11/01/2008 2:29:00 PM Breanna GALLAGHER:

No comment was recorded

Approver

Samantha CAVE

Approval/ Rejection Date

18/01/2008

Approval/ Rejection Comment

18/01/2008 12:00:00 AM Sam CAVE:

Approve action carried out, no changes to procedure required reminders to staff sufficient

Associations

No Associations found

Records

No Records found

18580 No Title Provided

Copyright © 2015, Health Services Support Agency, Queensland Health - All Rights Reserved